

WYLER AG, Im Hölderli CH-8405 Winterthur www.wylerag.com +41 (0) 52 233 66 66

## Declaration of Conformity EU DoC





WYLER AG declares hereby under the sole responsibility that the product

Name

Description
Part Numbers

Clinotronic PLUS

precise inclination maesurement instrument

015-PLUS-xxzz \*)

\*) xx := different Options like magnets and threads

yy := different measuring ranges like ±10°/ ±30°/ ±45°/ ±60°

has been submitted to a conformity assessment and fulfils the essential requirements of the following European directives

2014/30/EU

EMC (electro magnetic compatibility)

2011/65/ EU/ 2015/863/ EU

RohS II (restriction of hazardous substances)

and complies with the following harmonized standards or other normative documents

EN 61326-1: 2013

Electrical equipment for measurement, control and laboratory use,

IEC 61326-1: 2012

**EMC** requirements

EN 50581: 2012

Technical documentation for the assessment of electrical and

electronic products with respect to the restriction of

hazardous substances

The authorized person to compile the technical file and issue the declaration of conformity:

Name / Position

Heinz E. Hinnen / CEO

Signature

Place and date of issue

Winterthur, 27. Juli 2018

Author: P. Breu Verified: Rev. 2 / 12.09.2018 DoC\_Clinotronic PLUS\_R2.doc